

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Waiting Rm Room#: 101 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr
Department: IR Suite													
BKT-TVW		1	Bracket, Television, Wall, Flat Panel	Peerless-AV (SUF640P)	Project	Draft (New)	No	No	No	W	20.69	N/A	N/A
	O/V		DesignerSeries Univ. Ultra Slim (for	Peerless-AV (SUF640P)	Yes	1-Fixed	No	No	No	4	0.33	N/A	N/A
			37"- 50")		Yes	Unassigned	No	No	No	5	17.95	N/A	No
			Load capacity: 150 lbs. Metal stud fastener kit (sold separately)										
CRG-PHN		1	Charger, Telephone	ChargeTech (BIC)	Project	Draft (New)	No	No	No			N/A	N/A
	O/O		Phone Charging Station	ChargeTech (BIC)	No	2-Movable, Elect	No	No	No			N/A	N/A
					No	Unassigned	No	No	No			N/A	No
DSP-SZF		1	Dispenser, Hand Sanitizer,	GOJO Industries (2454-DS02)	Project	Draft (New)	No	No	No	F	16.12	N/A	N/A
	O/O		Freestanding	GOJO Industries (2454-DS02)	No	3-Movable, Non-	No	No	No	60	11.12	N/A	N/A
			Purell Elite Floor Stand (2454-DS02)		No	Elect	No	No	No	60	54.35	N/A	No
			Ships 2/case. Includes 3 "C" batteries for operation.										
TVM-49L		1	Television, 46-49 in., Flat Panel	LG Commercial Products	Project	Draft (New)	No	No	No	S	43.70	120	Type B (NEMA 5-15)
	O/O		LT340H Series 49" LED	(49LT340H0UA)	Yes	2-Movable, Elect	No	No	No	25	2.80	.8333	N/A
				LG Commercial Products	No	Unassigned	No	No	No	38	25.60	60	No
			Dimensions/weight reflect without stand. (Depth to Back VESA Location) Depth with stand: 3.2" Weight with stand: 32 lbs.										
WHC-AEL		1	Wheelchair, Adult, Extra Large	Medline Industries Inc.	Project	Draft (New)	No	No	No	M	30.50	N/A	N/A
	O/O		Excel Extra Wide - 22"	(MDS806800FLA)	No	3-Movable, Non-	No	No	No	88	40.00	N/A	N/A
			[MDS806800FLA]	Medline Industries Inc.	No	Elect	No	No	No	88	36.50	N/A	No
				(MDS806800FLA)		Unassigned							No
WST-SRD		1	Waste Can, Open Top	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	17.60	N/A	N/A
	O/O		Silhouettes Half Round 12 Gal	(FGSH12EPLSM)	No	3-Movable, Non-	No	No	No	22	8.80	N/A	N/A
			Perforated	Rubbermaid Commercial Products	No	Elect	No	No	No	26	32.40	N/A	No
				(FGSH12EPLSM)		Unassigned							No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Reception (2) Room#: 101A Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water		Vacuum		(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type		
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.	
Department: IR Suite															
CMP-AL1		2	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)		
	O/O		ThinkCentre M700z (20") All-In-One (Non-Touch)	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A Yes		
			Dimensions include stand.		No	Unassigned	No	No	No	18	16.25	60	No No		
PRC-CCD		2	Processor, Credit Card	First Data Corporation (FD200-TI)	Project	Draft (New)	No	No	No	C	5.25	120	Type B (NEMA 5-15)		
	O/V		FD200Ti	First Data Corporation (FD200-TI)	Yes	6-IT/Computers	No	No	No	4	8.00	1.5	N/A Yes		
			Dimensions are approximate, actual dimensions unavailable from manufacturer.		No	Unassigned	No	No	No	6	2.50	60	No No		
SCN-CTP		2	Scanner, Document, Countertop	Fujitsu (PA03750-B005)	Project	Draft (New)	No	No	No	C	11.50	120	Type B (NEMA 5-15)		
	O/O		fi-7030	Fujitsu (PA03750-B005)	No	2-Movable, Elect	No	No	No	7	5.75	0.1	N/A No		
					No	Unassigned	No	No	No	7	5.25	60	No No		
TEL-DSK		2	Telephone, Desktop	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A		
	O/V		TBD	TBD ()	Yes	6-IT/Computers	No	No	No			N/A	N/A No		
					No	Unassigned	No	No	No			N/A	No No		
WST-40Q		1	Waste Can, Open Top	Rubbermaid Commercial Products (FG254400BEIG)	Project	Draft (New)	No	No	No	F	11.30	N/A	N/A		
	O/O		2544 Fire Resistant Beige 40 qt	Rubbermaid Commercial Products (FG254400BEIG)	No	3-Movable, Non-Elect	No	No	No	6	15.00	N/A	N/A No		
				Rubbermaid Commercial Products (FG254400BEIG)	No	Unassigned	No	No	No	6	20.00	N/A	No No		

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Copier Room#: 101A; 113C Room Sign: Area/Phase: IR Suite Room Qty: 2

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
BIN-SHU		1	Bin, Shredding, Secure	Shred-it, Inc. ()	Project	Draft (New)	No	No	No	F	20.25	N/A	N/A	
	O/V		Mini Console	Shred-it, Inc. ()	No	3-Movable, Non-	No	No	No	30	19.50	N/A	N/A	No
			Fits under workstation.		No	Elect Unassigned	No	No	No	40	26.00	N/A	No	No
CPY-MFL		1	Copier, Floor, Multifunction	Ricoh Corporation (417046)	Project	Draft (New)	No	No	No	F	23.25	120	Type B (NEMA 5-15)	
	O/V		Aficio MP 5054	Ricoh Corporation (417046)	Yes	2-Movable, Elect	No	No	No	164	27.00	12	N/A	Yes
			Final weight and dimensions vary per customer and all options selected.			No	Unassigned	No	No	No	200	38.00	60	Yes
WST-16R		1	Waste Can, Recycle	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.50	N/A	N/A	
	O/O		Slim Jim Vented 16 Gal/Blue w/Vertical	(1971257/2018258)	No	3-Movable, Non-	No	No	No	8	22.00	N/A	N/A	No
			Lid	Rubbermaid Commercial Products	(1971257/2018258)	No	Elect Unassigned	No	No	No	8	31.00	N/A	No

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Exam Rm/Consult Room#: 104 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water			Vacuum		(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type			
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.		
Department: IR Suite																
BKT-TVC		1	Bracket, Television, Ceiling, Flat Panel	Peerless-AV (PLCM-UNL-CP)	Project	Draft (New)	No	No	No	CE	24.38	N/A	N/A			
	O/C		SmartMount Univ. w/Ceiling Mount (32"	Peerless-AV (PLCM-UNL-CP)	Yes	1-Fixed	No	No	No	26	4.03	N/A	N/A	No		
			- 90")		No	Unassigned	No	No	No	32	44.23	N/A	No	No		
			Load capacity: 150 lb.													
CMP-AL1		1	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)			
	O/O		ThinkCentre M700z (20") All-In-One	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A	Yes		
			(Non-Touch)		No	Unassigned	No	No	No	18	16.25	60	No	No		
			Dimensions include stand.													
DGS-IMN		1	Diagnostic System, Integrated,	Hillrom - Welch Allyn, Inc. (85NTVEC-B)	Project	Draft (New)	No	No	No	W	39.92	120	Type B (NEMA 5-15)			
	O/C		w/Monitoring		Yes	1-Fixed	No	No	No	14	7.51	1.5	N/A	Yes		
			Connex Integrated Wall System	Hillrom - Welch Allyn, Inc. (85NTVEC-B)	No	Unassigned	No	No	No	20	10.56	60	No	No		
			(Wireless)													
			Weight includes battery. Width and reflects with Braun Thermometer.													
DSL-S4G		1	Disposal, Sharps, Wall Mount	Stericycle (C-04RES-04/WB-04)	Project	Draft (New)	No	No	No	W	13.75	N/A	N/A			
	O/C		Bio Systems C-04RES-04 w/Locking	Stericycle (C-04RES-04/WB-04)	Yes	1-Fixed	No	No	No	13	7.75	N/A	N/A	No		
			Bracket		No	Unassigned	No	No	No	6	21.25	N/A	No	No		
			Max weight reflects container full.													
DSP-GV3		1	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A			
	O/C		305362 Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A	No		
					No	Unassigned	No	No	No	1	15.50	N/A	No	No		
DSP-PTA		1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A			
	O/C		Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No		
			enMotion Impulse 8" Automated		No	Unassigned	No	No	No	9	13.80	N/A	No	No		
			(White)													
			Recommended distance from bottom of bracket to the floor is 46 to 51 inches.													
			Mounting should comply with ADA guidelines. Operates on four D-cell alkaline													
			batteries.													
DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A			
	O/C		Provon NXT Maximum Capacity (2215-	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No		
			08)		No	Unassigned	No	No	No	12	10.65	N/A	No	No		
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware.													
			Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm)													
			clearance from bottom of dispenser to surface.													
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A			
	O/C		Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No		
			Mounts to wall with adhesive (included) or screws (not included). Requires 10"		No	Unassigned	No	No	No	7	10.51	N/A	No	No		
			(25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.													

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Exam Rm/Consult Room#: 104 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer Vendor	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model		Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
DSP-WPS		1	Dispenser, Disinfectant Wipes, Wall Mount	Metrex (13-1175)	Project	Draft (New)	No	No	No	W	5.25	N/A	N/A	
	O/C		CaviWipes Canister Bracket	Metrex (13-1175)	Yes	1-Fixed	No	No	No	1	5.25	N/A	N/A	No
			Ship weight for case of 12 is 15 lbs.		No	Unassigned	No	No	No	1	6.25	N/A	No	No
SCL-AMB		1	Scale, Clinical, Adult, Mechanical Beam	Detecto Scale- Div.Cardinal Scale Mfg (439)	Project	Draft (New)	No	No	No	F	10.50	N/A	N/A	
	O/O		439 Eye-Level, Lb	Detecto Scale- Div.Cardinal Scale Mfg (439)	No	3-Movable, Non-Elect	No	No	No	38	18.50	N/A	N/A	No
					No	Unassigned	No	No	No	38	59.00	N/A	No	No
STL-EBR		1	Stool, Exam, w/Backrest	Midmark Corporation (425-001/002-0273/002-0274)	Project	Draft (New)	No	No	No	M	23.00	N/A	N/A	
	O/O		Midmark 425 Air Lift w/Backrest	Midmark Corporation (425-001/002-0273/002-0274)	No	3-Movable, Non-Elect	No	No	No	20	23.00	N/A	N/A	No
					No	Unassigned	No	No	No	25	33.50	N/A	No	No
TBL-EXM		1	Table, Exam/Treatment, Manual Adjust, Electric	Midmark Corporation (204-002/ 002-0871)	Project	Draft (New)	No	No	No	F	30.00	115	Type B (NEMA 5-15)	
	O/O		Ritter 204 (Dwr Htr, Pelvic Tilt, Seamless Top)	Midmark Corporation (204-002/ 002-0871)	Yes	2-Movable, Elect	No	No	No	220	76.00	5.5	N/A	No
					Yes	Unassigned	No	No	No	300	33.13	60	No	No
TEL-DSK		1	Telephone, Desktop	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A	
	O/V		TBD	TBD ()	Yes	6-IT/Computers	No	No	No			N/A	N/A	No
					No	Unassigned	No	No	No			N/A	No	No
TVM-32H		1	Television, 30-32 in, Flat Panel	LG Commercial Products (32LT572MBUC)	Project	Draft (New)	No	No	No	S	29.10	120	Type B (NEMA 5-15)	
	O/O		32 in. LT572M Series Pro Centric Hospital	LG Commercial Products (32LT572MBUC)	Yes	2-Movable, Elect	No	No	No	11	3.30	0.3583	N/A	Yes
					No	Unassigned	No	No	No	13	17.40	60	No	No
WST-13B		1	Waste Can, Bio-Hazardous	Rubbermaid Commercial Products (1883566)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
	O/O		1883566 Slim Jim Resin Front Step 13 Gal Red	Rubbermaid Commercial Products (1883566)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
					No	Unassigned	No	No	No	9	28.30	N/A	No	No
WST-13W		1	Waste Can, Step-On	Rubbermaid Commercial Products (1883458)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
	O/O		Slim Jim Resin Front Step 13 Gal/Beige	Rubbermaid Commercial Products (1883458)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
					No	Unassigned	No	No	No	9	28.30	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail

Room: Exam Rm/Consult		Room#: 104	Room Sign:	Area/Phase: IR Suite	Room Qty: 1																
Comments:																					
Alt ID	Item ID	Qty	Description	Manufacturer Vendor	Funding Source		Item Status		Water		Vacuum		(lbs)	(in)	Electrical/Mechanical						
		F/I	Model		Arch Sig.	Spat. Sig.	Arch Code	Custom Code	Cold	Drain	Dental	Mounting	Max Weight	Depth	Volts	Plug Type					
		Spec Remarks / Item Notes							Hot	Steam	Medical	Ship Weight			Amp	Data Cnct.					
									Treated	Gas	Vent	Height			Hertz	BTU/hr	Emrg. Pwr.				
Department: IR Suite																					

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Procedure Rm Room#: 105 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water Cold Hot Treated	Vacuum Drain Steam Gas	Dental Medical Vent	(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical Volts Amp Hz	Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.
Department: IR Suite														
BKT-TVC	O/C	1	Bracket, Television, Ceiling, Flat Panel	Peerless-AV (PLCM-UNL-CP)	Project	Draft (New)	No	No	No	CE	24.38	N/A	N/A	
			SmartMount Univ. w/Ceiling Mount (32"	Peerless-AV (PLCM-UNL-CP)	Yes	1-Fixed	No	No	No	26	4.03	N/A	N/A	No
			- 90") Load capacity: 150 lb.		No	Unassigned	No	No	No	32	44.23	N/A	No	No
CMP-AL1	O/O	1	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)	
			ThinkCentre M700z (20") All-In-One	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A	Yes
			(Non-Touch) Dimensions include stand.		No	Unassigned	No	No	No	18	16.25	60	No	No
CRT-CYM	O/O	1	Cart, Cylinder, M, Single	Anthony Welded Products, Inc. (54PN)	Project	Draft (New)	No	No	No	M	22.00	N/A	N/A	
			54PN Heavy Duty	Anthony Welded Products, Inc. (54PN)	No	3-Movable, Non-Elect	No	No	No	37	22.00	N/A	N/A	No
					No	Unassigned	No	No	No	37	48.00	N/A	No	No
CRT-PGN	O/O	1	Cart, Procedure, General	InterMetro Industries Corp	Project	Draft (New)	No	No	No	M	22.50	N/A	N/A	
			Basix Plus Compact MBPC1210TL	(MBPC1210TL)	No	3-Movable, Non-Elect	No	No	No	133	24.25	N/A	N/A	No
				InterMetro Industries Corp ()	No	Unassigned	No	No	No	133	38.00	N/A	No	No
CRT-WOW	O/O	1	Cart, Computer, Workstation	JACO Inc. (220)	Project	Draft (New)	No	No	No	M	20.00	N/A	N/A	
			UltraLite 220 (LED/LCD mount)	JACO Inc. (220)	No	2-Movable, Elect	No	No	No	52	16.00	N/A	N/A	No
			Height adjusts 30" to 46"		No	Unassigned	No	No	No	65	46.00	N/A	No	No
DSL-S4G	O/C	1	Disposal, Sharps, Wall Mount	Stericycle (C-04RES-04/WB-04)	Project	Draft (New)	No	No	No	W	13.75	N/A	N/A	
			Bio Systems C-04RES-04 w/Locking	Stericycle (C-04RES-04/WB-04)	Yes	1-Fixed	No	No	No	13	7.75	N/A	N/A	No
			Bracket Max weight reflects container full.		No	Unassigned	No	No	No	6	21.25	N/A	No	No
DSP-GV3	O/C	1	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A	
			305362 Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A	No
					No	Unassigned	No	No	No	1	15.50	N/A	No	No
DSP-PTA	O/C	1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A	
			Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No
			enMotion Impulse 8" Automated (White) Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.		No	Unassigned	No	No	No	9	13.80	N/A	No	No

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Procedure Rm Room#: 105 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
DSP-SOP	O/C	1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A	
			Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.		No	Unassigned	No	No	No	12	10.65	N/A	No	No
DSP-STZ	O/C	1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
			Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No
HMP-MED	O/O	1	Hamper, Linen	Centurion Medical Products (CX302)	Project	Draft (New)	No	No	No	M	18.50	N/A	N/A	
			CX302 (Medium)	Centurion Medical Products (CX302)	No	3-Movable, Non-Elect	No	No	No	21	19.50	N/A	N/A	No
					No	Unassigned	No	No	No	21	32.75	N/A	No	No
LGT-EMB	O/O	1	Light, Exam/Procedure, Single, Mobile,	Getinge Group - MAQUET Surgical Workflows (ARDLCA309009A)	Project	Draft (New)	No	No	No	M	35.40	120	Type B (NEMA 5-15)	
			Articulating Arm	Getinge Group - MAQUET Surgical Workflows (ARDLCA309009A)	No	2-Movable, Elect	No	No	No	21	35.40	0.5	N/A	No
			Lucea LED 40 Mobile	Getinge Group - MAQUET Surgical Workflows (ARDLCA309009A)	No	Unassigned	No	No	No	34	79.90	60	No	No
MON-PHB	O/O	1	Monitor, Physiologic, Bedside	Philips Healthcare - Monitoring Systems (865241-A06)	Project	Draft (New)	No	No	No	S	15.50	120	Type B (NEMA 5-15)	
			IntelliVue MX700 (General/ICU)	Philips Healthcare - Monitoring Systems (865241-A06)	Yes	2-Movable, Elect	No	No	No	21	6.50	1.7	N/A	Yes
				Philips Healthcare - Monitoring Systems (865241-A06)	No	Unassigned	No	No	No	40	12.75	60	No	Yes
SND-IVP	O/O	1	Stand, IV, Chrome	Medline Industries Inc. (MDS80494)	Project	Draft (New)	No	No	No	M	24.00	N/A	N/A	
			Deluxe IV [MDS80494]	Medline Industries Inc. (MDS80494)	No	3-Movable, Non-Elect	No	No	No	9	24.00	N/A	N/A	No
					No	Unassigned	No	No	No	12	96.00	N/A	No	No
STD-EMN	O/O	1	Stand, Equipment, Monitor	Philips Healthcare - Monitoring Systems (989803137481)	Project	Draft (New)	No	No	No	M	25.00	N/A	N/A	
			989803137481	Philips Healthcare - Monitoring Systems ()	No	3-Movable, Non-Elect	No	No	No	25	25.00	N/A	N/A	No
					No	Unassigned	No	No	No	30	43.00	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Procedure Rm Room#: 105 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
STR-PRC		1	Stretcher, Procedure / Recovery w/ IV pole	Stryker Medical (1115-000-030)	Project	Draft (New)	No	No	No	M	38.00	N/A	N/A	
	O/O		Prime with Big Wheel 1115 (30' Litter)	Stryker Medical (1115-000-030)	No	3-Movable, Non-Elect	No	No	No	342	85.00	N/A	N/A	No
					Yes	Unassigned	No	No	No	385	48.00	N/A	No	No
TVM-32H		1	Television, 30-32 in, Flat Panel	LG Commercial Products (32LT572MBUC)	Project	Draft (New)	No	No	No	S	29.10	120	Type B (NEMA 5-15)	
	O/O		32 in. LT572M Series Pro Centric Hospital	LG Commercial Products (32LT572MBUC)	Yes	2-Movable, Elect	No	No	No	11	3.30	0.3583	N/A	Yes
					No	Unassigned	No	No	No	13	17.40	60	No	No
WST-13B		1	Waste Can, Bio-Hazardous	Rubbermaid Commercial Products (1883566)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
	O/O		1883566 Slim Jim Resin Front Step 13 Gal Red	Rubbermaid Commercial Products (1883566)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
					No	Unassigned	No	No	No	9	28.30	N/A	No	No
WST-13W		1	Waste Can, Step-On	Rubbermaid Commercial Products (1883458)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
	O/O		Slim Jim Resin Front Step 13 Gal/Beige	Rubbermaid Commercial Products (1883458)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
					No	Unassigned	No	No	No	9	28.30	N/A	No	No

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Toilet/Patient Room#: 106 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
BAR-GB2		3	Bar, Grab, Toilet/Shower	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No	W	26.00	N/A	N/A	
		O/C	B-6806 x 24 Straight Grab Bar	6806 x 24)	Yes	1-Fixed	No	No	No	2	3.00	N/A	N/A	No
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No	2	3.25	N/A	No	No
			Dimensions reflect bar mounted horizontally. Clearance between the grab bar and wall 1-1/2 inches.											
DSP-PTT		1	Dispenser, Paper Towel, with Trash	Georgia Pacific (59466A/59491)	Project	Draft (New)	No	No	No	W	17.25	N/A	N/A	
		O/C	Receptacle	Georgia Pacific (59466A/59491)	Yes	1-Fixed	No	No	No	36	8.00	N/A	N/A	No
			enMotion 8" Recessed w/Receptacle		No	Unassigned	No	No	No	50	56.00	N/A	No	No
			(Stainless)											
DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A	
		O/C	Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
					No	Unassigned	No	No	No	12	10.65	N/A	No	No
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.											
DSP-STN		1	Dispenser, Sanitary Napkin	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No			N/A	N/A	
		O/C	B-35139 TrimLine Series	370634 25)	Yes	1-Fixed	No	No	No			N/A	N/A	No
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No			N/A	No	No
				370634 25)										
DSP-STN		1	Dispenser, Sanitary Napkin	Bobrick Washroom Equipment, Inc. ()	Project	Draft (New)	No	No	No			N/A	N/A	
		O/C	B-35139 TrimLine Series	Bobrick Washroom Equipment, Inc. ()	Yes	1-Fixed	No	No	No			N/A	N/A	No
					No	Unassigned	No	No	No			N/A	No	No
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
		O/C	Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
			Mounts to wall with adhesive (included) or screws (not included). Requires 10"		No	Unassigned	No	No	No	7	10.51	N/A	No	No
			(25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.											
DSP-TPS		1	Dispenser, Toilet Paper, Surface	Bobrick Washroom Equipment, Inc. ()	Project	Draft (New)	No	No	No			N/A	N/A	
		O/C	Mount	Bobrick Washroom Equipment, Inc. ()	Yes	1-Fixed	No	No	No			N/A	N/A	No
			B-3588-Multi Trim Line Series		No	Unassigned	No	No	No			N/A	No	No
DSP-TSC		1	Dispenser, Toilet Seat Cover	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No	W	15.75	N/A	N/A	
		O/C	B-221 ClassicSeries Surface Mount	221)	Yes	1-Fixed	No	No	No	3	2.00	N/A	N/A	No
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No	3	11.00	N/A	No	No
				221)										

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Toilet/Patient **Room#:** 106 **Room Sign:** **Area/Phase:** IR Suite **Room Qty:** 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Item Status		Water		Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
					Arch Sig.	Spat. Sig.	Arch Code	Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent				Volts Amp Hz	Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.

Department: IR Suite

Recommended Mounting Height Off Floor 45in. to 59in. (1145 to 1500mm)

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: EVS Room#: 108 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer Vendor	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model		Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
		Spec	Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
BUK-HYG		1	Bucket, Mopping	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	23.88	N/A	N/A	
	O/O	HYGEN	Microfiber Charging Bucket	(FGQ95088YEL)	No	3-Movable, Non-Elect	No	No	No	5	9.50	N/A	N/A	No
				Rubbermaid Commercial Products (FGQ95088YEL)	No	Unassigned	No	No	No	5	13.63	N/A	No	No
DSP-CNS		1	Dispenser, Cleaning Solution	Diversey Care (D04379)	Project	Draft (New)	Yes	No	No	W	10.00	N/A	N/A	
	O/C	J-Fill Duo		Diversey Care (D04379)	Yes	1-Fixed	No	No	No	25	6.50	N/A	N/A	No
					No	Unassigned	No	No	No	32	19.50	N/A	No	No
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
	O/C	Purell FMX-12/1000-1200ml	(5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
		Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.			No	Unassigned	No	No	No	7	10.51	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Soiled Work Rm Room#: 109 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr
Department: IR Suite													
DSL-S4G		1	Disposal, Sharps, Wall Mount	Stericycle (C-04RES-04/WB-04)	Project	Draft (New)	No	No	No	W	13.75	N/A	N/A
	O/C		Bio Systems C-04RES-04 w/Locking Bracket	Stericycle (C-04RES-04/WB-04)	Yes	1-Fixed	No	No	No	13	7.75	N/A	N/A
			Max weight reflects container full.		No	Unassigned	No	No	No	6	21.25	N/A	No
DSP-GV3		1	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A
	O/C		305362 Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A
					No	Unassigned	No	No	No	1	15.50	N/A	No
DSP-PTA		1	Dispenser, Paper Towel, Surface Mount	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A
	O/C		enMotion Impulse 8" Automated (White)	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A
			Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.		No	Unassigned	No	No	No	9	13.80	N/A	No
DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A
	O/C		Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.		No	Unassigned	No	No	No	12	10.65	N/A	No
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A
	O/C		Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No
HMP-SML		1	Hamper, Linen	Centurion Medical Products (CX502)	Project	Draft (New)	No	No	No	M	18.50	N/A	N/A
	O/O		CX502 (Small)	Centurion Medical Products (CX502)	No	3-Movable, Non-Elect	No	No	No	17	19.50	N/A	N/A
					No	Unassigned	No	No	No	17	25.75	N/A	No
WST-24B		1	Waste Can, Bio-Hazardous	Rubbermaid Commercial Products (1883570)	Project	Draft (New)	No	No	No	F	22.40	N/A	N/A
	O/O		1883570 Slim Jim Resin Front Step 24 Gal/Red	Rubbermaid Commercial Products (1883570)	No	3-Movable, Non-Elect	No	No	No	13	13.90	N/A	N/A
					No	Unassigned	No	No	No	13	32.50	N/A	No
WST-24W		1	Waste Can, Step-On	Rubbermaid Commercial Products (1883552)	Project	Draft (New)	No	No	No	F	22.40	N/A	N/A
	O/O		1883552 Slim Jim Resin Front Step 24 Gal Beige	Rubbermaid Commercial Products (1883552)	No	3-Movable, Non-Elect	No	No	No	13	13.90	N/A	N/A
					No	Unassigned	No	No	No	13	32.50	N/A	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Soiled Work Rm Room#: 109 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical											
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type									
							Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.								
							Treated	Gas	Vent	Ship Weight	Height	Hz	Ded. Circ.	Emrg. Pwr.								
Spec Remarks / Item Notes															Spat. Sig.	Custom Code						

Department: IR Suite

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Lockers Patient Room#: 109A Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
							Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
							Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
							Gas	Vent	Ship Weight	Height	Hz	Ded. Circ.	Emrg. Pwr.	
Department: IR Suite														
	LKR-Z1H	3	Locker, 1 Tier	Hollman ()	Project	Draft (New)	No	No	No	F	12.00	N/A	N/A	
		O/C	Z1	Hollman ()	Yes	1-Fixed	No	No	No	44	18.00	N/A	N/A	No
			Height includes 6 inch legs. Bolts to wall and floor.	No	Unassigned	No	No	No	55	78.00	N/A	No	No	

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Pre/Post Bay 01,02,03 Room#: 110; 111; 112 Room Sign: Area/Phase: IR Suite Room Qty: 3

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical											
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type									
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.								
														Treated	Gas	Vent	Ship Weight	Height	Hz	Ded. Circ.	Emrg. Pwr.	
Department: IR Suite																						
BKT-CWS		1	Bracket, Computer Workstation	Humanscale Healthcare (VF56-0505-14032)	Project	Draft (New)	No	No	No	W	28.00	N/A	N/A									
		O/C	V/Flex Wall Station w/CPU Holder	Humanscale Healthcare (VF56-0505-14032)	Yes	1-Fixed	No	No	No	45	41.00	N/A	N/A	No								
					No	Unassigned	No	No	No	45	56.00	N/A	No	No								
			Provide electrical outlet for CPU, which mounts separately to wall. Weight is estimated and not provided by supplier. Width reflects keyboard tray. Depth reflects max arm extension. Height reflects track.																			
BKT-TVC		1	Bracket, Television, Ceiling, Flat Panel	Peerless-AV (PLCM-UNL-CP)	Project	Draft (New)	No	No	No	CE	24.38	N/A	N/A									
		O/C	SmartMount Univ. w/Ceiling Mount (32"	Peerless-AV (PLCM-UNL-CP)	Yes	1-Fixed	No	No	No	26	4.03	N/A	N/A	No								
			- 90")		No	Unassigned	No	No	No	32	44.23	N/A	No	No								
			Load capacity: 150 lb.																			
CMP-AL1		1	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)									
		O/O	ThinkCentre M700z (20") All-In-One	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A	Yes								
			(Non-Touch)		No	Unassigned	No	No	No	18	16.25	60	No	No								
			Dimensions include stand.																			
CRT-CYM		1	Cart, Cylinder, M, Single	Anthony Welded Products, Inc. (54PN)	Project	Draft (New)	No	No	No	M	22.00	N/A	N/A									
		O/O	54PN Heavy Duty	Anthony Welded Products, Inc. (54PN)	No	3-Movable, Non-Elect	No	No	No	37	22.00	N/A	N/A	No								
					No	Unassigned	No	No	No	37	48.00	N/A	No	No								
CRT-PGN		1	Cart, Procedure, General	InterMetro Industries Corp	Project	Draft (New)	No	No	No	M	22.50	N/A	N/A									
		O/O	Basix Plus Compact MBPC1210TL	(MBPC1210TL)	No	3-Movable, Non-Elect	No	No	No	133	24.25	N/A	N/A	No								
				InterMetro Industries Corp ()	No	Unassigned	No	No	No	133	38.00	N/A	No	No								
DSL-S4G		1	Disposal, Sharps, Wall Mount	Stericycle (C-04RES-04/WB-04)	Project	Draft (New)	No	No	No	W	13.75	N/A	N/A									
		O/C	Bio Systems C-04RES-04 w/Locking Bracket	Stericycle (C-04RES-04/WB-04)	Yes	1-Fixed	No	No	No	13	7.75	N/A	N/A	No								
			Max weight reflects container full.		No	Unassigned	No	No	No	6	21.25	N/A	No	No								
DSP-GV3		1	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A									
		O/C	305362 Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A	No								
					No	Unassigned	No	No	No	1	15.50	N/A	No	No								
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A									
		O/C	Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No								
			Mounts to wall with adhesive (included) or screws (not included). Requires 10"		No	Unassigned	No	No	No	7	10.51	N/A	No	No								
			(25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.																			

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Pre/Post Bay 01,02,03 Room#: 110; 111; 112 Room Sign: Area/Phase: IR Suite Room Qty: 3

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	Data Cnct.
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Emrg. Pwr.
Department: IR Suite														
MON-PHB		1	Monitor, Physiologic, Bedside	Philips Healthcare - Monitoring	Project	Draft (New)	No	No	No	S	15.50	120	Type B (NEMA 5-15)	
	O/O		IntelliVue MX700 (General/ICU)	Systems (865241-A06)	Yes	2-Movable, Elect	No	No	No	21	6.50	1.7	N/A	Yes
				Philips Healthcare - Monitoring	No	Unassigned	No	No	No	40	12.75	60	No	Yes
			Specifications are for MX700 monitor only. FMS-4 module rack is 9"W x 5.5"H x 7.5"D/ 2.4 lbs. Mounts to wall bracket or rolling stand.											
SND-IVP		1	Stand, IV, Chrome	Medline Industries Inc. (MDS80494)	Project	Draft (New)	No	No	No	M	24.00	N/A	N/A	
	O/O		Deluxe IV [MDS80494]	Medline Industries Inc. (MDS80494)	No	3-Movable, Non-Elect	No	No	No	9	24.00	N/A	N/A	No
					No	Unassigned	No	No	No	12	96.00	N/A	No	No
STD-EMN		1	Stand, Equipment, Monitor	Philips Healthcare - Monitoring	Project	Draft (New)	No	No	No	M	25.00	N/A	N/A	
	O/O		989803137481	Systems (989803137481)	No	3-Movable, Non-Elect	No	No	No	25	25.00	N/A	N/A	No
				Philips Healthcare - Monitoring	No	Unassigned	No	No	No	30	43.00	N/A	No	No
STR-PRC		1	Stretcher, Procedure / Recovery w/ IV	Stryker Medical (1115-000-030)	Project	Draft (New)	No	No	No	M	38.00	N/A	N/A	
	O/O		pole	Stryker Medical (1115-000-030)	No	3-Movable, Non-Elect	No	No	No	342	85.00	N/A	N/A	No
			Prime with Big Wheel 1115 (30' Litter)		Yes	Unassigned	No	No	No	385	48.00	N/A	No	No
TVM-32H		1	Television, 30-32 in, Flat Panel	LG Commercial Products	Project	Draft (New)	No	No	No	S	29.10	120	Type B (NEMA 5-15)	
	O/O		32 in. LT572M Series Pro Centric	(32LT572MBUC)	Yes	2-Movable, Elect	No	No	No	11	3.30	0.3583	N/A	Yes
			Hospital	LG Commercial Products	No	Unassigned	No	No	No	13	17.40	60	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Sink Handwash Room#: 110A; 112A; 113B Room Sign: Area/Phase: IR Suite Room Qty: 3

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
DSP-PTA	O/C	1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A	
			Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No
			enMotion Impulse 8" Automated (White)		No	Unassigned	No	No	No	9	13.80	N/A	No	No
			Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.											
DSP-SOP	O/C	1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A	
			Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.		No	Unassigned	No	No	No	12	10.65	N/A	No	No
DSP-STZ	O/C	1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
			Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No
WST-13W	O/O	1	Waste Can, Step-On	Rubbermaid Commercial Products (1883458)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
			Slim Jim Resin Front Step 13	Rubbermaid Commercial Products (1883458)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
			Gal/Beige		No	Unassigned	No	No	No	9	28.30	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Equipment Shared Room#: 110B;112B Room Sign: Area/Phase: IR Suite Room Qty: 2

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water		Vacuum		(lbs)	(in)	Electrical/Mechanical									
		F/I	Model				Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type						
		Spec	Remarks / Item Notes				Spat. Sig.	Custom Code	Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.						
															Treated	Gas	Vent	Ship Weight	Height	Hz	Ded. Circ.	Emrg. Pwr.
Department: IR Suite																						
HMP-MED		1	Hamper, Linen	Centurion Medical Products (CX302)	Project	Draft (New)	No	No	No	M	18.50	N/A	N/A									
		O/O	CX302 (Medium)	Centurion Medical Products (CX302)	No	3-Movable, Non-Elect	No	No	No	21	19.50	N/A	N/A	No								
					No	Unassigned	No	No	No	21	32.75	N/A	No	No								
WST-13B		1	Waste Can, Bio-Hazardous	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A									
		O/O	1883566 Slim Jim Resin Front Step 13	(1883566)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No								
			Gal Red	Rubbermaid Commercial Products (1883566)	No	Unassigned	No	No	No	9	28.30	N/A	No	No								
WST-13W		1	Waste Can, Step-On	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A									
		O/O	Slim Jim Resin Front Step 13	(1883458)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No								
			Gal/Beige	Rubbermaid Commercial Products (1883458)	No	Unassigned	No	No	No	9	28.30	N/A	No	No								

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Nurse Station (1) Room#: 113 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water Cold Hot Treated	Vacuum Drain Steam Gas	Dental Medical Vent	(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
												Volts	Plug Type	Data Cnct.
												Amp Hz	BTU/hr Ded. Circ.	Emrg. Pwr.
Department: IR Suite														
CMP-AL1		1	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)	
		O/O	ThinkCentre M700z (20") All-In-One (Non-Touch)	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A	Yes
			Dimensions include stand.		No	Unassigned	No	No	No	18	16.25	60	No	No
SCN-CTP		1	Scanner, Document, Countertop	Fujitsu (PA03750-B005)	Project	Draft (New)	No	No	No	C	11.50	120	Type B (NEMA 5-15)	
		O/O	fi-7030	Fujitsu (PA03750-B005)	No	2-Movable, Elect	No	No	No	7	5.75	0.1	N/A	No
					No	Unassigned	No	No	No	7	5.25	60	No	No
TEL-DSK		1	Telephone, Desktop	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A	
		O/V	TBD	TBD ()	Yes	6-IT/Computers	No	No	No			N/A	N/A	No
					No	Unassigned	No	No	No			N/A	No	No
WST-40Q		1	Waste Can, Open Top	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.30	N/A	N/A	
		O/O	2544 Fire Resistant Beige 40 qt	(FG254400BEIG)	No	3-Movable, Non-Elect	No	No	No	6	15.00	N/A	N/A	No
				Rubbermaid Commercial Products (FG254400BEIG)	No	Unassigned	No	No	No	6	20.00	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Crash Cart-Warming Cabinet Room#: 113A Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
CAB-W1G		1	Cabinet, Warming, Single,	MAC Medical, Inc. (SWC242464-G-4B)	Project	Draft (New)	No	No	No	F	24.00	120	Type B (NEMA 5-15)	
		O/O	Freestanding	MAC Medical, Inc. (SWC242464-G-4B)	Yes	2-Movable, Elect	No	No	No	315	26.50	6.5	1604	Yes
			D-Series SWC242464-G-4B (Glass Door)		No	Unassigned	No	No	No	315	64.75	60	Yes	No
			Includes 6 ft. power cord, located at top/rear of unit. 15 amp dedicated circuit recommended.											
CRT-CDE		1	Cart, Procedure, Resuscitation	Armstrong Medical Industries (ABL-AR-6)	Project	Draft (New)	No	No	No	M	34.25	N/A	N/A	
		O/O	ABL-AR-6 Standard Steel Breakaway	Armstrong Medical Industries (ABL-AR-6)	No	3-Movable, Non-	No	No	No	165	25.50	N/A	N/A	No
			Lock 6-Dwr (Beige/Red)	Armstrong Medical Industries (ABL-AR-6)	No	Elect	No	No	No	175	45.50	N/A	No	No
						Unassigned								
DFB-MNP		1	Defibrillator, Monitor, w/Pacing	Zoll Medical Corporation (3 0120 0090 0131 0012)	Project	Draft (New)	No	No	No	C	10.50	120	Type B (NEMA 5-15)	
		O/O	R Series ALS w/Exp Pkg, Masimo SpO2, NIBP & Pacing	Zoll Medical Corporation (3 0120 0090 0131 0012)	Yes	2-Movable, Elect	No	No	No	15	12.50	N/A	N/A	No
					No	Unassigned	No	No	No	20	8.20	60	No	Yes
			Electrical amps not available, contact Zoll for more information.											
PMP-SAG		1	Pump, Suction/Aspirator, General,	Armstrong Medical Industries (AE-6976)	Project	Draft (New)	No	No	No	S	17.00	120	Type B (NEMA 5-15)	
		O/O	Portable	Armstrong Medical Industries (AE-6976)	No	2-Movable, Elect	No	No	No	11	6.00	<1	N/A	No
			SSCOR DUET w/Retention Bracket	Armstrong Medical Industries (AE-6976)	No	Unassigned	No	No	No	12	9.00	60	No	No
			Mounts to wall or cart or other device. Electrical requirements are for battery charger.											
WST-13W		1	Waste Can, Step-On	Rubbermaid Commercial Products (1883458)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A	
		O/O	Slim Jim Resin Front Step 13	Rubbermaid Commercial Products (1883458)	No	3-Movable, Non-	No	No	No	9	11.50	N/A	N/A	No
			Gal/Beige	Rubbermaid Commercial Products (1883458)	No	Elect	No	No	No	9	28.30	N/A	No	No
						Unassigned								

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: Meds Room#: 113D Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
DSP-GV3		1	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A	
	O/C	305362	Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A	No
					No	Unassigned	No	No	No	1	15.50	N/A	No	No
DSP-MDM		1	Dispenser, Medication, Host (Main)	Omnicell, Inc. (MED-FRM-104)	Project	Draft (New)	No	No	No	F	26.50	115	Type B (NEMA 5-15)	
	O/V		Omnicell XT Half-Height Cabinet	Omnicell, Inc. (MED-FRM-104)	Yes	2-Movable, Elect	No	No	No	275	27.00	3	400	Yes
			Optional External Return Bin adds to overall width and requires clearance for door opening. Max Weight reflects empty, actual weight is varies by configuration. Units must be within 6 feet of electrical outlet. Power and data connections are made to frames at back right of computer.		No	Unassigned	No	No	No	275	53.30	60	No	Yes
DSP-MLK		1	Dispenser, Medication, Lock Module	Omnicell, Inc. (SRD-OPT-033)	Project	Draft (New)	No	No	No	S	5.50	115	Type B (NEMA 5-15)	
	O/O		Omnicell XT FlexLock Wireless	Omnicell, Inc. (SRD-OPT-033)	Yes	2-Movable, Elect	No	No	No	8	12.40	0.8	N/A	Yes
			Installs on refrigerator (sold separately from third party). Refrigerator specs vary. Allow 2 in. side clearance for key access. Requires installation of a Bluetooth dongle on the main cabinet. (Bluetooth is optional feature implemented with Omnicell products starting with software version 21.5.) Maximum distance for pairing FlexLock and main cabinet should 50 ft. (15.24 m), line of sight.		No	Unassigned	No	No	No	8	5.50	60	No	No
DSP-PTA		1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A	
	O/C		Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No
			enMotion Impulse 8" Automated (White)		No	Unassigned	No	No	No	9	13.80	N/A	No	No
			Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.											
DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A	
	O/C		Provon NXT Maximum Capacity (2215- GOJO Industries (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.		No	Unassigned	No	No	No	12	10.65	N/A	No	No
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
	O/C		Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No
MON-GLC		1	Monitor, Blood Glucose, Point-of-Care	Roche Diagnostics Corporation	Project	Draft (New)	No	No	No	C	4.75	120	Type B (NEMA 5-15)	
	O/O		ACCU-CHEK Inform II Meter & Base	(05060311001/05060290001)	No	2-Movable, Elect	No	No	No	2	4.00	1.7	N/A	Yes
			Unit	Roche Diagnostics Corporation (05060311001/05060290001)	No	Unassigned	No	No	No	4	7.75	60	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Meds Room#: 113D Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water		Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
							Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz			Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
Department: IR Suite															
REF-PUC		1	Refrigerator, Pharmaceutical, Undercounter PH-ABT-HC-UCBI-0404G-ADA Premier (4.6 cu.ft) Front clearance for door swing. Built-in model has front breathing ventilation system. Unit should not be connected to a GFI outlet.	American BioTech Supply (PH-ABT-HC-UCBI-0404G-ADA) American BioTech Supply (PH-ABT-HC-UCBI-0404G-ADA)	Project	Draft (New)	No	No	No	F	23.75	115	Type B (NEMA 5-15)		
		O/C			Yes	2-Movable, Elect	No	No	No	97	25.50	1.74	N/A		
					No	Unassigned	No	No	No	150	31.94	60	No		
WST-13W		1	Waste Can, Step-On Slim Jim Resin Front Step 13 Gal/Beige	Rubbermaid Commercial Products (1883458) Rubbermaid Commercial Products (1883458)	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A		
		O/O			No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A		
					No	Unassigned	No	No	No	9	28.30	N/A	No		

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Workstations (2) Room#: 114 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water Cold Hot Treated	Vacuum Drain Steam Gas	Dental Medical Vent	(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
												Volts	Plug Type	Data Cnct.
												Amp Hz	BTU/hr Ded. Circ.	Emrg. Pwr.
Department: IR Suite														
CMP-AL1		2	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)	
		O/O	ThinkCentre M700z (20") All-In-One (Non-Touch)	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A	Yes
			Dimensions include stand.		No	Unassigned	No	No	No	18	16.25	60	No	No
SCN-CTP		2	Scanner, Document, Countertop	Fujitsu (PA03750-B005)	Project	Draft (New)	No	No	No	C	11.50	120	Type B (NEMA 5-15)	
		O/O	fi-7030	Fujitsu (PA03750-B005)	No	2-Movable, Elect	No	No	No	7	5.75	0.1	N/A	No
					No	Unassigned	No	No	No	7	5.25	60	No	No
TEL-DSK		2	Telephone, Desktop	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A	
		O/V	TBD	TBD ()	Yes	6-IT/Computers	No	No	No			N/A	N/A	No
					No	Unassigned	No	No	No			N/A	No	No
WST-40Q		2	Waste Can, Open Top	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.30	N/A	N/A	
		O/O	2544 Fire Resistant Beige 40 qt	(FG254400BEIG)	No	3-Movable, Non-Elect	No	No	No	6	15.00	N/A	N/A	No
				Rubbermaid Commercial Products (FG254400BEIG)	No	Unassigned	No	No	No	6	20.00	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Lounge/Staff Room#: 115 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical				
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type		
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.	
Department: IR Suite															
	LKR-Z1H	4	Locker, 2 Tier	Hollman ()	Project	Draft (New)	No	No	No	F	12.00	N/A	N/A		
		O/C	Z1	Hollman ()	Yes	1-Fixed	No	No	No	45	18.00	N/A	N/A	No	
			Height includes 6 inch legs. Bolts to wall and floor.		No	Unassigned	No	No	No	56	78.00	N/A	No	No	
CLK-AWL		1	Clock, Analog, Synchronized, Wireless	La Crosse Technology (WT-3122A)	Project	Draft (New)	No	No	No	W	12.50	N/A	N/A		
		O/O	WT-3122A 12.5 inch Atomic Wall Clock	La Crosse Technology (WT-3122A)	Yes	1-Fixed	No	No	No	3	1.50	N/A	N/A	No	
			Requires 1 AA battery. Depth from Manufacturer website. Weight from vendor.		No	Unassigned	No	No	No	3	12.50	N/A	No	No	
CLK-EMP		1	Clock, Employee Time	KRONOS Incorporated (9100)	Project	Draft (New)	No	No	No	W	10.75	120	Type B (NEMA 5-15)		
		O/O	InTouch 9100	KRONOS Incorporated (9100)	No	2-Movable, Elect	No	No	No	5	4.00	1.5	N/A	Yes	
					No	Unassigned	No	No	No	5	6.00	60	No	No	
COF-SGC		1	Coffee Maker, Single Cup, Plumbed	Keurig Green Mountain (K150P)	Project	Draft (New)	Yes	No	No	C	10.40	120	Type B (NEMA 5-15)		
		O/O	K150P Commercial Brewing System - Plumbed	Keurig Green Mountain (K150P)	Yes	1-Fixed	No	No	No	18	14.00	11.7	N/A	No	
			Specifications from Manufacturer product webpage. Weight reflects empty brewing system. Manufacturer recommends 15 amp dedicated, grounded circuit. Requires the use of an external water filter such as Omnipure KQ8. Sold seperately. Does not require drain. Drain hose in back of unit - water drains into pitcher.		No	Unassigned	No	No	No	22	13.90	60	Yes	No	
DSP-PTA		1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A		
		O/C	Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No	
			enMotion Impulse 8" Automated (White)		No	Unassigned	No	No	No	9	13.80	N/A	No	No	
			Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.												
		DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A
				O/C	Provon NXT Maximum Capacity (2215- GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
	Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.				No	Unassigned	No	No	No	12	10.65	N/A	No	No	
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A		
		O/C	Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No	
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No	

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Lounge/Staff Room#: 115 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
OVN-MCW		1	Oven, Domestic, Microwave,	GE Appliances (JEM3072DHWW)	Project	Draft (New)	No	No	No	C	17.25	120	Type B (NEMA 5-15)	
	O/O		Countertop	GE Appliances (JEM3072DHWW)	Yes	2-Movable, Elect	No	No	No	21	13.00	9.5	N/A	No
			0.7 cu. ft. Capacity, White		Yes	Unassigned	No	No	No	24	10.13	60	No	No
Microwave must be grounded. Depth with door open = 25 5/8 in														
REF-UCA		1	Refrigerator, Domestic, Undercounter	Summit Appliance (AL750W)	Project	Draft (New)	No	No	No	F	23.63	115	Type B (NEMA 5-15)	
	O/O		AL750W (ADA Compliant)	Summit Appliance (AL750W)	Yes	2-Movable, Elect	No	No	No	100	24.00	0.9	N/A	No
			Depth with handle shown. Depth with door at 90 degrees 45.13".		No	Unassigned	No	No	No	105	32.25	60	Yes	No
Separate circuit recommended. 15 amps circuit required.														
TEL-WMT		1	Telephone, Wall	TBD ()	Project	Draft (New)	No	No	No	W	N/A	N/A	N/A	
	O/V		TBD	TBD ()	Yes	6-IT/Computers	No	No	No	N/A	N/A	N/A	N/A	Yes
					No	Unassigned	No	No	No	N/A	N/A	N/A	No	No
WST-16R		1	Waste Can, Recycle	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.50	N/A	N/A	
	O/O		Slim Jim Vented 16 Gal/Blue w/Vertical	(1971257/2018258)	No	3-Movable, Non-	No	No	No	8	22.00	N/A	N/A	No
			Lid	Rubbermaid Commercial Products	No	Elect Unassigned	No	No	No	8	31.00	N/A	No	No
(1971257/2018258)														
WST-24W		1	Waste Can, Step-On	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	22.40	N/A	N/A	
	O/O		1883552 Slim Jim Resin Front Step 24	(1883552)	No	3-Movable, Non-	No	No	No	13	13.90	N/A	N/A	No
			Gal Beige	Rubbermaid Commercial Products	No	Elect Unassigned	No	No	No	13	32.50	N/A	No	No
(1883552)														

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Toilet/Staff Room#: 116 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water		Vacuum		(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type		
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam Gas	Medical Vent	Max Weight Ship Weight	Depth Height	Amp Hz	BTU/hr Ded.	Data Cnct. Emrg. Pwr.	
Department: IR Suite															
BAR-GB2		3	Bar, Grab, Toilet/Shower	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No	W	26.00	N/A	N/A		
	O/C		B-6806 x 24 Straight Grab Bar	6806 x 24)	Yes	1-Fixed	No	No	No	2	3.00	N/A	N/A	No	
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No	2	3.25	N/A	No	No	
			Dimensions reflect bar mounted horizontally. Clearance between the grab bar and wall 1-1/2 inches.												
DSP-PTT		1	Dispenser, Paper Towel, with Trash	Georgia Pacific (59466A/59491)	Project	Draft (New)	No	No	No	W	17.25	N/A	N/A		
	O/C		Receptacle	Georgia Pacific (59466A/59491)	Yes	1-Fixed	No	No	No	36	8.00	N/A	N/A	No	
			enMotion 8" Recessed w/Receptacle (Stainless)		No	Unassigned	No	No	No	50	56.00	N/A	No	No	
DSP-SOP		1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A		
	O/C		Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No	
					No	Unassigned	No	No	No	12	10.65	N/A	No	No	
			Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.												
DSP-STN		1	Dispenser, Sanitary Napkin	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No			N/A	N/A		
	O/C		B-35139 TrimLine Series	370634 25)	Yes	1-Fixed	No	No	No			N/A	N/A	No	
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No			N/A	No	No	
DSP-STN		1	Dispenser, Sanitary Napkin	Bobrick Washroom Equipment, Inc. ()	Project	Draft (New)	No	No	No			N/A	N/A		
	O/C		B-35139 TrimLine Series	Bobrick Washroom Equipment, Inc. ()	Yes	1-Fixed	No	No	No			N/A	N/A	No	
					No	Unassigned	No	No	No			N/A	No	No	
DSP-STZ		1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A		
	O/C		Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No	
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No	
DSP-TPS		1	Dispenser, Toilet Paper, Surface	Bobrick Washroom Equipment, Inc. ()	Project	Draft (New)	No	No	No			N/A	N/A		
	O/C		Mount	Bobrick Washroom Equipment, Inc. ()	Yes	1-Fixed	No	No	No			N/A	N/A	No	
			B-3588-Multi Trim Line Series		No	Unassigned	No	No	No			N/A	No	No	
DSP-TSC		1	Dispenser, Toilet Seat Cover	Bobrick Washroom Equipment, Inc. (B-	Project	Draft (New)	No	No	No	W	15.75	N/A	N/A		
	O/C		B-221 ClassicSeries Surface Mount	221)	Yes	1-Fixed	No	No	No	3	2.00	N/A	N/A	No	
				Bobrick Washroom Equipment, Inc. (B-	No	Unassigned	No	No	No	3	11.00	N/A	No	No	

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Toilet/Staff Room#: 116 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source			Item Status			Water			Vacuum			(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
					Arch Sig.	Spat. Sig.	Arch Code	Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent								Volts Amp Hz	Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.

Department: IR Suite

Recommended Mounting Height Off Floor 45in. to 59in. (1145 to 1500mm)

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: IR Procedure Rm Room#: 117 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer Vendor	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical		
		F/I	Model		Arch Sig.	Arch Code	Cold Hot Treated	Drain Steam	Dental Medical	Mounting Max Weight	Width Data	Volts	Plug Type
		Spec Remarks / Item Notes	Spat. Sig.		Custom Code		Gas	Vent	Ship Weight	Depth Height	Amp Hz	BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.
Department: IR Suite													
APR-LGW		10	Apron, Lead	Wolf X-Ray Corporation (65023)	Project	Draft (New)	No	No	No	S	42.00	N/A	N/A
		O/O	65023 Easywrap Large	Wolf X-Ray Corporation (65023)	No	3-Movable, Non-Elect	No	No	No	8	1.00	N/A	N/A
					No	Unassigned	No	No	No	10	24.00	N/A	No
CLK-AWL		1	Clock, Analog, Synchronized, Wireless	La Crosse Technology (WT-3122A)	Project	Draft (New)	No	No	No	W	12.50	N/A	N/A
		O/O	WT-3122A 12.5 inch Atomic Wall Clock	La Crosse Technology (WT-3122A)	Yes	1-Fixed	No	No	No	3	1.50	N/A	N/A
			Requires 1 AA battery. Depth from Manufacturer website. Weight from vendor.		No	Unassigned	No	No	No	3	12.50	N/A	No
CMP-AL1		2	Computer, Desktop, All-in-One	Lenovo (M700z)	Project	Draft (New)	No	No	No	C	20.00	120	Type B (NEMA 5-15)
		O/O	ThinkCentre M700z (20") All-In-One (Non-Touch)	Lenovo (M700z)	Yes	6-IT/Computers	No	No	No	15	9.75	1	N/A
			Dimensions include stand.		No	Unassigned	No	No	No	18	16.25	60	No
CRT-CYM		1	Cart, Cylinder, M, Single	Anthony Welded Products, Inc. (54PN)	Project	Draft (New)	No	No	No	M	22.00	N/A	N/A
		O/O	54PN Heavy Duty	Anthony Welded Products, Inc. (54PN)	No	3-Movable, Non-Elect	No	No	No	37	22.00	N/A	N/A
					No	Unassigned	No	No	No	37	48.00	N/A	No
CRT-ETS		1	Cart, Cylinder, D&E, Single	Anthony Welded Products, Inc. (6105)	Project	Draft (New)	No	No	No	M	12.00	N/A	N/A
		O/O	6105 Gurney Hook w/ 6" Wheels	Anthony Welded Products, Inc. (6105)	No	3-Movable, Non-Elect	No	No	No	7	10.00	N/A	N/A
					No	Unassigned	No	No	No	7	41.00	N/A	No
CRT-SCT		6	Cart, Supply, Enclosed w/ Catheter	Innerspace - Solaire	Project	Draft (New)	No	No	No	M	39.00	N/A	N/A
		O/O	Storage Assembly	(SR2G/STD8/STD4/STD2/S27S1/S9C M)	No	3-Movable, Non-Elect	No	No	No	193	28.75	N/A	N/A
			Roam 2 Supply w/Glass Dr, Trays, Shelf, Cath Mngrs	Innerspace - Solaire (SR2G/STD8/STD4/STD2/S27S1/S9C M)	No	Unassigned	No	No	No	193	75.25	N/A	No
CRT-US3		1	Cart, Utility, Stainless	Lakeside Manufacturing, Inc. (322)	Project	Draft (New)	No	No	No	M	18.38	N/A	N/A
		O/O	322 (3 Shelf, Standard Duty)	Lakeside Manufacturing, Inc. (322)	No	3-Movable, Non-Elect	No	No	No	32	30.75	N/A	N/A
					No	Unassigned	No	No	No	32	33.00	N/A	No
CRT-WOW		2	Cart, Computer, Workstation	JACO Inc. (220)	Project	Draft (New)	No	No	No	M	20.00	N/A	N/A
		O/O	UltraLite 220 (LED/LCD mount)	JACO Inc. (220)	No	2-Movable, Elect	No	No	No	52	16.00	N/A	N/A
			Height adjusts 30" to 46"		No	Unassigned	No	No	No	65	46.00	N/A	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: IR Procedure Rm Room#: 117 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
DSL-P8D		1	Disposal, Sharps, Floor Cart,	Stericycle (DWS-08/C-08BLKHAZ-PH)	Project	Draft (New)	No	No	No	M	14.25	N/A	N/A	
	O/O		Pharmacy	Stericycle (DWS-08/C-08BLKHAZ-PH)	No	3-Movable, Non-	No	No	No	12	16.50	N/A	N/A	
			Bio Systems DWS-08 w/C-08BLKHAZ-PH Black		No	Elect Unassigned	No	No	No	14	23.00	N/A	No	No
DSL-PHW		1	Disposal, Sharps, Wall Mount,	Stericycle (C-02BLKHAZ-PH /	Project	Draft (New)	No	No	No	W	13.00	N/A	N/A	
	O/C		Pharmacy	0203WMA)	Yes	1-Fixed	No	No	No	10	5.75	N/A	N/A	
			Bio Systems C-02BLKHAZ-PH /	Stericycle (C-02BLKHAZ-PH /	No	Unassigned	No	No	No	4	15.25	N/A	No	No
			0203WMA	0203WMA)										
			Max weight reflects container full. Width and Depth include Locking Wall Mount											
DSL-S4G		1	Disposal, Sharps, Wall Mount	Stericycle (C-04RES-04/WB-04)	Project	Draft (New)	No	No	No	W	13.75	N/A	N/A	
	O/C		Bio Systems C-04RES-04 w/Locking	Stericycle (C-04RES-04/WB-04)	Yes	1-Fixed	No	No	No	13	7.75	N/A	N/A	
			Bracket		No	Unassigned	No	No	No	6	21.25	N/A	No	No
			Max weight reflects container full.											
DSP-GV3		2	Dispenser, Glove, Triple Box	Omnimed, Inc (305362)	Project	Draft (New)	No	No	No	W	11.25	N/A	N/A	
	O/C		305362 Clear PETG	Omnimed, Inc (305362)	Yes	1-Fixed	No	No	No	1	3.75	N/A	N/A	
					No	Unassigned	No	No	No	1	15.50	N/A	No	No
DSP-SFL		2	Disposal, Sharps, Floor Cart	Stericycle (C-17/DWSH-17/LH-17)	Project	Draft (New)	No	No	No	M	14.50	N/A	N/A	
	O/O		Bio Systems C-17/DWSH-17/LH-17	Stericycle (C-17/DWSH-17/LH-17)	No	3-Movable, Non-	No	No	No	15	22.50	N/A	N/A	
					No	Elect Unassigned	No	No	No	18	29.75	N/A	No	No
DSP-STZ		2	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
	O/C		Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No
DSP-WPS		1	Dispenser, Disinfectant Wipes, Wall	Diversey Care (9P700826)	Project	Draft (New)	No	No	No	W	6.00	N/A	N/A	
	O/C		Mount	Diversey Care (9P700826)	Yes	1-Fixed	No	No	No	3	6.00	N/A	N/A	
			Oxivir Tb Wall Mount Wire Bracket		No	Unassigned	No	No	No	3	7.00	N/A	No	No
HMP-MED		1	Hamper, Linen	Centurion Medical Products (CX302)	Project	Draft (New)	No	No	No	M	18.50	N/A	N/A	
	O/O		CX302 (Medium)	Centurion Medical Products (CX302)	No	3-Movable, Non-	No	No	No	21	19.50	N/A	N/A	
					No	Elect Unassigned	No	No	No	21	32.75	N/A	No	No

Medical Equipment Resources, Ltd.

CUIMC-Interventional Radiology Suite

Specification Room by Room Detail



Room: IR Procedure Rm Room#: 117 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	Data Cnct.
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Emrg. Pwr.
Department: IR Suite														
IMG-CRM	O/V	1	X-Ray Unit, C-Arm, Mobile	Siemens Medical Imaging ()	Project	Draft (New)	No	No	No	M	32.90	120	Type B (NEMA 5-20)	
			Cios Spin	Siemens Medical Imaging ()	Yes	2-Movable, Elect	No	No	No	1391	94.50	20.00	N/A	Yes
			100 V to 127 V 20 A slow-blow fuse. Dimensions reflects C-arm chassis without options and monitor cart including 2 monitor, UPS. C-arm chasses depth ranges from 86.6 to 94.5 in. and height adjusts from 71 to 89 in. Monitor Cart: 29.1 x 27.6 x 70.7 to 76.9 in.		No	Unassigned	No	No	No	1391	89.00	60	No	No
IMG-UMP	O/O	1	Ultrasound, Imaging, Cardiac / Echo /	Siemens Medical Ultrasound	Project	Draft (New)	No	No	No	M	20.00	120	Type B (NEMA 5-15)	
			EchoPac S/W	(11361954)	Yes	2-Movable, Elect	No	No	No	168	33.25	8.3	2400	Yes
			Acuson Juniper	Siemens Medical Ultrasound (11361954)	No	Unassigned	No	No	No	168	64.50	60	No	No
Adjustable Depth 25.7 inches - 33.2 inches max. Adjustable height: (upright FPD) 48.2 – 64.3 inches.														
INJ-MBL	O/O	1	Injector, Contrast Media, Mobile	Bayer HealthCare Radiology (ART 700	Project	Draft (New)	No	No	No	M	26.10	120	Type B (NEMA 5-15)	
			Medrad Mark 7 Arterion	PEDL)	Yes	2-Movable, Elect	No	No	No	146	47.30	8.33	N/A	No
				Bayer HealthCare Radiology (ART 700 PEDL)	No	Unassigned	No	No	No	146	57.40	60	No	No
System includes mobile pedestal with injector head, integrated power unit and separate control display. See cutsheet for Room Planning details.														
LGT-EPC	O/C	1	Light, Exam/Procedure, Single, Ceiling	Getinge Group - MAQUET Surgical	Project	Draft (New)	No	No	No	CE	73.50	120	Hardwire	
			Lucea LED 40 (1200mm Suspension Tube)	Workflows (ARDLCA209021A)	Yes	1-Fixed	No	No	No	34	73.50	0.5	N/A	No
				Getinge Group - MAQUET Surgical Workflows (ARDLCA209021A)	No	Unassigned	No	No	No	34	ADJ	60	No	No
Width and depth reflect swivel range.														
MDG-AIR	O/O	1	Flowmeter, Air	Precision Medical (1MFA2005PTO)	Project	Draft (New)	No	No	No	W	1.25	N/A	N/A	
			Chrome (0-15 lpm, Ohmeda, Power Take Off)	Precision Medical (1MFA2005PTO)	No	3-Movable, Non-Elect	No	No	No	1	2.50	N/A	N/A	No
					No	Unassigned	No	Yes	No	1	6.50	N/A	No	No
MDG-OXF	O/O	2	Flowmeter, Oxygen	Precision Medical (1MFA1003PTO)	Project	Draft (New)	No	No	No	W	2.25	N/A	N/A	
			Chrome (0-15 lpm, DISS Female, PTO)	Precision Medical (1MFA1003PTO)	No	3-Movable, Non-Elect	No	No	No	1	2.50	N/A	N/A	No
					No	Unassigned	No	Yes	No	1	6.00	N/A	No	No
MDG-SCI	O/O	5	Regulator, Suction,	Precision Medical (PM3303)	Project	Draft (New)	No	No	No	W	2.80	N/A	N/A	
			Intermittent/Continuous	Precision Medical (PM3303)	No	3-Movable, Non-Elect	No	No	Yes	1	4.40	N/A	N/A	No
			PM3303 (DISS HT/Tubing Npl)		No	Unassigned	No	No	No	1	5.30	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: IR Procedure Rm Room#: 117 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
MON-PHB		1	Monitor, Physiologic, Bedside	Philips Healthcare - Monitoring	Project	Draft (New)	No	No	No	S	15.50	120	Type B (NEMA 5-15)	
	O/O		IntelliVue MX700 (General/ICU)	Systems (865241-A06)	Yes	2-Movable, Elect	No	No	No	21	6.50	1.7	N/A	Yes
				Philips Healthcare - Monitoring	No	Unassigned	No	No	No	40	12.75	60	No	Yes
			Specifications are for MX700 monitor only. FMS-4 module rack is 9"W x 5.5"H x 7.5"D/ 2.4 lbs. Mounts to wall bracket or rolling stand.											
RCK-AP5		2	Rack, Apron, Wall Mount	Infab Corporation (683100)	Project	Draft (New)	No	No	No	W	20.00	N/A	N/A	
	O/C		683100 (5-arm R/H folding)	Infab Corporation (683100)	Yes	1-Fixed	No	No	No	10	27.00	N/A	N/A	No
			Backing plate included from manufacturer. Mounting must accommodate weight of up to 5 lead aprons.		No	Unassigned	No	No	No	15	8.25	N/A	No	No
SHD-LDM		1	Shield, Lead, Mobile	AADCO Medical Inc. (S-601)	Project	Draft (New)	No	No	No	M	72.00	N/A	N/A	
	O/O		RayShield Ultra S-601 (72in.)	AADCO Medical Inc. (S-601)	No	3-Movable, Non-Elect	No	No	No	334	24.00	N/A	N/A	No
					Yes	Unassigned	No	No	No	334	75.00	N/A	No	No
SND-IVP		2	Stand, IV, Chrome	Medline Industries Inc. (MDS80494)	Project	Draft (New)	No	No	No	M	24.00	N/A	N/A	
	O/O		Deluxe IV [MDS80494]	Medline Industries Inc. (MDS80494)	No	3-Movable, Non-Elect	No	No	No	9	24.00	N/A	N/A	No
					No	Unassigned	No	No	No	12	96.00	N/A	No	No
STD-EMN		1	Stand, Equipment, Monitor	Philips Healthcare - Monitoring	Project	Draft (New)	No	No	No	M	25.00	N/A	N/A	
	O/O		989803137481	Systems (989803137481)	No	3-Movable, Non-Elect	No	No	No	25	25.00	N/A	N/A	No
				Philips Healthcare - Monitoring	No	Unassigned	No	No	No	30	43.00	N/A	No	No
STL-EBR		2	Stool, Exam, w/Backrest	Midmark Corporation (425-001/002-0273/002-0274)	Project	Draft (New)	No	No	No	M	23.00	N/A	N/A	
	O/O		Midmark 425 Air Lift w/Backrest	Midmark Corporation (425-001/002-0273/002-0274)	No	3-Movable, Non-Elect	No	No	No	20	23.00	N/A	N/A	No
					No	Unassigned	No	No	No	25	33.50	N/A	No	No
STP-HRL		1	Stool, Step, w/Handrail	Blickman Industries (1017757100)	Project	Draft (New)	No	No	No	F	18.00	N/A	N/A	
	O/O		7757MR-HR	Blickman Industries (1017757100)	No	3-Movable, Non-Elect	No	No	No	20	12.00	N/A	N/A	No
			Height reflects with handrail		No	Unassigned	No	No	No	20	41.00	N/A	No	No
TBL-IN6		2	Table, Instrument, 72 inch	Blickman Industries (0117837003)	Project	Draft (New)	No	No	No	M	72.00	N/A	N/A	
	O/O		7837SS-GR	Blickman Industries (0117837003)	No	3-Movable, Non-Elect	No	No	No	141	24.00	N/A	N/A	No
					Yes	Unassigned	No	No	No	141	34.00	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: IR Procedure Rm Room#: 117 Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical		
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr
Department: IR Suite													
TBL-SUG		1	Table, Surgical, Major	Skytron (3603H)	Project	Draft (New)	No	No	No	M	22.00	120	Type B (NEMA 5-15)
	O/O	3603H UltraSlide	Skytron (3603H)	Yes	2-Movable, Elect	No	No	No	904	82.00	15	N/A	No
		Height adjusts 23 to 41.5 inches. Table is equipped with a removable 15 ft. power cord for battery charging.		Yes	Unassigned	No	No	No	1130	41.50	60	No	No
TEL-DSK		1	Telephone, Desktop	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A
	O/V	TBD	TBD ()	Yes	6-IT/Computers	No	No	No			N/A	N/A	No
				No	Unassigned	No	No	No			N/A	No	No
THB-CTH		1	Thrombectomy System, Catheter	Boston Scientific (105650-TAB)	Project	Draft (New)	No	No	No	M	15.00	120	Type B (NEMA 5-15)
	O/O	AngioJet Ultra	Boston Scientific ()	No	2-Movable, Elect	No	No	No	150	23.00	10	N/A	No
				No	Unassigned	No	No	No	150	52.50	60	No	No
WMR-GL3		1	Warmer, Gel	Parker Laboratories Inc. (82-03)	Project	Draft (New)	No	No	No	C	7.25	120	Type B (NEMA 5-15)
	O/O	Thermasonic 82-03 LED (3-Bottle)	Parker Laboratories Inc. (82-03)	No	2-Movable, Elect	No	No	No	3	4.50	0.26	N/A	No
		Warmer can also be wall mounted.		No	Unassigned	No	No	No	5	8.50	60	No	No
WST-13B		1	Waste Can, Bio-Hazardous	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A
	O/O	1883566 Slim Jim Resin Front Step 13 Gal Red	(1883566)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
			Rubbermaid Commercial Products (1883566)	No	Unassigned	No	No	No	9	28.30	N/A	No	No
WST-13W		2	Waste Can, Step-On	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	17.90	N/A	N/A
	O/O	Slim Jim Resin Front Step 13 Gal/Beige	(1883458)	No	3-Movable, Non-Elect	No	No	No	9	11.50	N/A	N/A	No
			Rubbermaid Commercial Products (1883458)	No	Unassigned	No	No	No	9	28.30	N/A	No	No
WST-55D		1	Waste Can, 44-55 Gallon	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	M	26.40	N/A	N/A
	O/O	2655 BRUTE Gray w/2640 Dolly	(FG265500GRAY/FG264000BLA)	No	3-Movable, Non-Elect	No	No	No	22	26.40	N/A	N/A	No
			Rubbermaid Commercial Products (FG265500GRAY/FG264000BLA)	No	Unassigned	No	No	No	22	39.60	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Alcove/Scrub Room#: 117A Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical			
		F/I	Model	Vendor	Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type	
			Spec Remarks / Item Notes		Spat. Sig.	Custom Code	Hot	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.
Department: IR Suite														
DSP-PTA	O/C	1	Dispenser, Paper Towel, Surface	Georgia Pacific (59437A)	Project	Draft (New)	No	No	No	W	12.70	N/A	N/A	
			Mount	Georgia Pacific (59437A)	Yes	1-Fixed	No	No	No	6	8.58	N/A	N/A	No
			enMotion Impulse 8" Automated (White)		No	Unassigned	No	No	No	9	13.80	N/A	No	No
		Recommended distance from bottom of bracket to the floor is 46 to 51 inches. Mounting should comply with ADA guidelines. Operates on four D-cell alkaline batteries.												
DSP-SOP	O/C	1	Dispenser, Soap, Wall Mount	GOJO Industries (2215-08)	Project	Draft (New)	No	No	No	W	6.59	N/A	N/A	
			Provon NXT Maximum Capacity (2215-08)	GOJO Industries (2215-08)	Yes	1-Fixed	No	No	No	2	4.76	N/A	N/A	No
					No	Unassigned	No	No	No	12	10.65	N/A	No	No
		Ships 8/case. Mounts to wall with included adhesive tape or optional hardware. Multiple hole pattern allows use of existing wall holes. Requires 4" (10 cm) clearance from bottom of dispenser to surface.												
DSP-STZ	O/C	1	Dispenser, Hand Sanitizer, Wall Mount	GOJO Industries (5120-06)	Project	Draft (New)	No	No	No	W	5.98	N/A	N/A	
			Purell FMX-12/1000-1200ml (5120-06)	GOJO Industries (5120-06)	Yes	1-Fixed	No	No	No	1	3.98	N/A	N/A	No
			Mounts to wall with adhesive (included) or screws (not included). Requires 10" (25.4 cm) clearance from bottom of dispenser to surface. Ships 6/case.		No	Unassigned	No	No	No	7	10.51	N/A	No	No
SNK-S1B	O/C	1	Sink, Scrub, 1-Bay, Stainless Steel	Getinge Group - Getinge Infection Control (GSS101ENNWNA)	Project	Draft (New)	Yes	Yes	No	W	32.00	120	Hardwire	
			GSS101E (Electric, Wall Mount Bracket)	Getinge Group - Getinge Infection Control (GSS101ENNWNA)	Yes	1-Fixed	Yes	No	No	118	25.00	1	N/A	No
					No	Unassigned	No	No	No	148	55.25	60	No	No
		Per manufacturer, fused disconnect switch must be provided at sink location. Includes wall mount bracket kit. For use with existing concrete or reinforced walls.												
WST-24W	O/O	1	Waste Can, Step-On	Rubbermaid Commercial Products (1883552)	Project	Draft (New)	No	No	No	F	22.40	N/A	N/A	
			1883552 Slim Jim Resin Front Step 24 Gal Beige	Rubbermaid Commercial Products (1883552)	No	3-Movable, Non-Elect	No	No	No	13	13.90	N/A	N/A	No
					No	Unassigned	No	No	No	13	32.50	N/A	No	No

Medical Equipment Resources, Ltd.
CUIMC-Interventional Radiology Suite
Specification Room by Room Detail



Room: Miscellaneous Costs Room#: XXX Room Sign: Area/Phase: IR Suite Room Qty: 1

Comments:

Alt ID	Item ID	Qty	Description	Manufacturer Vendor	Funding Source	Item Status	Water	Vacuum	(lbs)	(in)	Electrical/Mechanical							
		F/I	Model		Arch Sig.	Arch Code	Cold	Drain	Dental	Mounting	Width	Volts	Plug Type					
		Spec	Remarks / Item Notes		Spat. Sig.	Custom Code	Hot Treated	Steam	Medical	Max Weight	Depth	Amp	BTU/hr	Data Cnct.				
														Ship Weight	Height	Hz	Ded. Circ.	Emrg. Pwr.
Department: IR Suite																		
ZZZ-FGT		1	Services, Freight/Shipping	TBD ()	Project	Draft (New)	No	No	No	N/A	N/A	N/A	N/A					
		O/V	TBD	TBD ()	No	0-Unassigned	No	No	No	N/A	N/A	N/A	N/A	No				
					No	Unassigned	No	No	No	N/A	N/A	N/A	No	No				
ZZZ-INL		1	Services, Installation	TBD ()	Project	Draft (New)	No	No	No			N/A	N/A					
		O/V	TBD	TBD ()	No	0-Unassigned	No	No	No	N/A	N/A	N/A	N/A	No				
					No	Unassigned	No	No	No	N/A	N/A	N/A	No	No				
ZZZ-SPG		10	Allowance, Paging	TBD (TBD)	Project	Draft (New)	No	No	No	N/A	N/A	N/A	N/A					
		O/O	TBD	TBD (TBD)	Yes	1-Fixed	No	No	No	N/A	N/A	N/A	N/A	No				
					No	Unassigned	No	No	No	N/A	N/A	N/A	No	No				
ZZZ-WRH		1	Services, Storage/Warehousing	TBD (TBD)	Project	Draft (New)	No	No	No	N/A	N/A	N/A	N/A					
		O/V	TBD	TBD (TBD)	No	0-Unassigned	No	No	No	N/A	N/A	N/A	N/A	No				
					No	Unassigned	No	No	No	N/A	N/A	N/A	No	No				